

# The MINERvA Operations Report

## All Experimenters Meeting

Howard Budd, University of Rochester  
Sep 10, 2012





# MINERvA Detector



- Detector was powered up after replacing the UPS which had failed during the July 10 power outage
  - UPS is heavy (~ 100 lb), and we needed a procedure and equipment to replace the UPS.
- Running shifts starting 5:30PM weekdays for about 1½ hours
  - We are leaving HV off except for shifts
  - Run PEDs and Light Injection to determine if there has been any change in the detector.
    - There appears to be no change in the detector or PMTs
  - Shifter posts plots of 3 sets vibration sensors
    - Geophones on PMTs
    - Seismometers at north end of MINOS Hall
    - Water level sensors